



# Declaration of Conformity

<b>PRODUCT IDENTIFICATION</b>		
<b>Product name</b>	<b>Model/number</b>	
VBMax Pulmonary Function Filter	1563xx 1565xx 18xxxx	1564xx 1568xx
<b>MANUFACTURER</b>		
<b>Name of company</b>	<b>Address</b>	<b>Representative</b>
A-M Systems, Inc.	PO Box 850 Carlsborg, WA 98324 USA	Rob Thompson
<b>AUTHORIZED REPRESENTATIVE</b>		
<b>Name of company</b>	<b>Address</b>	<b>Telephone/email</b>
Emergo Europe	Molenstraat 15 2513 BH The Hague, Netherlands	+31.70.345.8570 - phone +31.70.346.7299 - fax service@emergogroup.com
<b>CONFORMITY ASSESSMENT</b>		
<b>Device classification</b>	<b>Route to compliance</b>	<b>Standards applied</b>
Class IIa Rule 2	Annex V of MDD 93/42/EEC Council Directive	
<b>REGISTRATION INFORMATION</b>		
<b>Notified Body and ID #</b>	<b>CE certificate number</b>	
BSI, 0086	CE 533831	

A-M Systems, Inc. declares that the above-mentioned products meet the provision of the Council Directive 93/42/EEC for Medical Devices.

**COMPANY REPRESENTATIVE:** Mr. Arthur B. Green III

**TITLE:** General Manager    **SIGNATURE:**

**DATE:** 14/12/2007