



Declaration of Conformity

PRODUCT IDENTIFICATION		
Product name	Model/number	
ViroMax Viral & Bacterial Filter	1562xx	
PRODUCT SPECIFICATIONS		
Specification Name	Value	
Bacterial Filtration Efficiency	>99.999%	
Viral Filtration Efficiency	>99.99%	
MANUFACTURER		
Name of company	Address	Representative
A-M Systems, LLC	131 Business Park Loop Sequim, WA 98382 USA	Robert Thompson
AUTHORIZED REPRESENTATIVE		
Name of company	Address	Telephone/email
Emergo Europe	Prinsessegracht 20 2514 AP The Hague The Netherlands	+31.70.345.8570 - phone +31.70.346.7299 - fax europe@emergogroup.com
CONFORMITY ASSESSMENT		
Device classification	Route to compliance	Standards applied
Class IIa Rule 2	Annex V of MDD 93/42/EEC Council Directive	ISO 14971 ISO 13485 ISO 15223-1 BS EN 1041 ISO 5356-1
REGISTRATION INFORMATION		
Notified Body and ID #	CE certificate number	
BSI, 0086	CE 533831	

A-M Systems declares that the above mentioned products meet the provision of the Council Directive 93/42/EEC for Medical Devices and Directive 93/42/EEC as transposed in the national laws of the Member States.

COMPANY REPRESENTATIVE: Mr. Robert Thompson

TITLE: Quality Manager **SIGNATURE:** 

DATE: 05/06/2017

P.O. Box 850
Carlsborg, WA 98324
U.S.A.

World Wide Web: www.a-msystems.com
Customer Service E-Mail: sales@a-msystems.com
Technical Support E-Mail: tech@a-msystems.com

Tel: 800-426-1306
Tel: 360-683-8300
Fax: 360-683-3525